
Midmark M11 - v1.0.5
Total Cycles: 3372

Name: RITTER2
Sterilizer ID: V2313228
02/09/2024 11:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.0 F 0.0
2:00 143.5 F 0.8
4:00 149.2 F 3.4
6:00 174.3 F 7.1
8:00 196.0 F 11.5
10:00 214.5 F 5.2
12:00 223.3 F 10.1
14:00 231.0 F 8.1
16:00 241.4 F 12.9
18:00 242.1 F 12.0
20:00 254.9 F 18.7
22:00 266.3 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 271.6 F 29.6
2:00 272.2 F 29.6
3:00 271.3 F 29.0
4:00 271.6 F 29.8

Min 271.0 F 29.0
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.6 F 0.0
5:00 212.9 F 0.0
10:00 214.8 F 0.0
15:00 209.8 F 0.0
20:00 208.4 F 0.0
25:00 206.6 F 0.0
30:00 206.0 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 23:31
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:01:52
CYCLE COMPLETE

02/09/2024 12:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____