
Midmark M11 - v1.0.5
Total Cycles: 2548

Name: RITTER2
Sterilizer ID: V2313228
06/21/2023 02:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.3 F 0.0
2:00 194.2 F 4.1
4:00 203.4 F 9.3
6:00 219.5 F 7.4
8:00 229.6 F 10.4
10:00 232.7 F 7.9
12:00 241.1 F 11.2
14:00 246.7 F 13.8
16:00 259.2 F 20.5
18:00 269.9 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.6 F 29.4
2:00 271.8 F 29.2
3:00 272.4 F 29.4
4:00 271.7 F 29.0

Min 271.7 F 28.9
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 228.0 F 0.0
10:00 235.6 F 0.0
15:00 227.6 F 0.0
20:00 227.3 F 0.0
25:00 228.1 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 2:26
HEATING: 19:43
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:59:29
CYCLE COMPLETE

06/21/2023 03:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____