
Midmark M11 - v1.0.5
Total Cycles: 2443

Name: RITTER2
Sterilizer ID: V2313228
05/31/2023 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.3 F 0.0
2:00 199.7 F 4.3
4:00 209.5 F 10.2
6:00 223.7 F 6.4
8:00 236.7 F 12.4
10:00 237.8 F 10.5
12:00 237.0 F 9.4
14:00 251.9 F 16.6
16:00 263.9 F 23.8
18:00 271.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.6
2:00 271.8 F 28.9
3:00 271.9 F 29.4
4:00 272.5 F 29.5

Min 271.3 F 28.7
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 227.7 F 0.0
10:00 239.9 F 0.0
15:00 234.2 F 0.0
20:00 225.8 F 0.0
25:00 225.0 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 2:28
HEATING: 18:01
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:39
CYCLE COMPLETE

05/31/2023 11:48 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____