

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2606

Name: RITTER 1  
Sterilizer ID: V2178198  
11/23/2022 03:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.7 F 0.0  
2:00 206.3 F 5.1  
4:00 214.1 F 10.9  
6:00 226.4 F 6.7  
8:00 236.5 F 12.9  
10:00 238.8 F 10.9  
12:00 238.5 F 9.7  
14:00 253.5 F 17.2  
16:00 266.0 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.6 F 29.1  
2:00 271.7 F 28.4  
3:00 272.1 F 29.1  
4:00 272.6 F 29.0  
  
Min 271.4 F 28.2  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.1 F 0.0  
5:00 229.4 F 0.0  
10:00 240.3 F 0.0  
15:00 231.2 F 0.0  
20:00 230.4 F 0.0  
25:00 230.8 F 0.0  
30:00 225.1 F 0.0  
  
Drying Complete

FILLING: 2:15  
HEATING: 17:34  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:57:01  
CYCLE COMPLETE

11/23/2022 04:11 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_