

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4616

Name: RITTER 1  
Sterilizer ID: V2178198  
06/07/2024 09:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.5 F 0.0  
2:00 180.7 F 2.0  
4:00 195.1 F 5.5  
6:00 208.6 F 9.7  
8:00 226.2 F 8.8  
10:00 229.2 F 7.7  
12:00 239.0 F 12.1  
14:00 235.5 F 8.8  
16:00 241.9 F 10.9  
18:00 244.8 F 12.7  
20:00 256.3 F 18.7  
22:00 266.5 F 25.1  
24:00 272.9 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.3 F 28.9  
2:00 272.7 F 29.2  
3:00 271.4 F 28.9  
4:00 272.2 F 29.0

Min 271.4 F 28.7  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.1 F 0.0  
5:00 236.4 F 0.0  
10:00 215.2 F 0.0  
15:00 231.5 F 0.0  
20:00 223.8 F 0.0  
25:00 228.5 F 0.0  
30:00 235.8 F 0.0  
Drying Complete

FILLING: 1:30  
HEATING: 24:22  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:03:39  
CYCLE COMPLETE

06/07/2024 10:22 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_