

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002569  
7:31 10/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:50  
PRESSURIZING 2:56  
STERILIZING 4:03  
136.4 C 223kPa 4:03  
136.4 C 225kPa 4:33  
136.2 C 223kPa 5:03  
136.4 C 224kPa 5:33  
136.2 C 224kPa 6:03  
136.4 C 224kPa 6:33  
136.4 C 223kPa 7:03  
136.0 C 221kPa 7:33  
136.0 C 221kPa 7:33  
VENTING 7:33  
AIR DRYING 8:23  
CYCLE COMPLETE 9:34  
-----

Digital Signature #  
F5858DC355F716C7  
-----