
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001483
9:26 26/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 3:48
STERILIZING 4:52
136.0 C 221kPa 4:52
136.4 C 222kPa 5:22
136.2 C 222kPa 5:52
136.2 C 222kPa 6:22
136.3 C 224kPa 6:52
136.4 C 223kPa 7:22
136.5 C 221kPa 7:52
136.2 C 222kPa 8:22
136.2 C 222kPa 8:22
VENTING 8:22
AIR DRYING 9:14
CYCLE COMPLETE 9:20

Digital Signature #
DCE27D23EB6CC0C3
