
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002301
8:49 27/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:11
STERILIZING 4:23
136.4 C 220kPa 4:23
136.1 C 222kPa 4:53
136.2 C 223kPa 5:23
136.3 C 221kPa 5:53
136.4 C 223kPa 6:23
136.2 C 223kPa 6:53
136.3 C 223kPa 7:23
136.4 C 223kPa 7:53
136.4 C 223kPa 7:53
VENTING 7:53
AIR DRYING 8:46
CYCLE COMPLETE 8:48

Digital Signature #
1E72635EA1CCC0F5
