

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3033

Name: RITTER2  
Sterilizer ID: V2313228  
10/19/2023 01:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.3 F 0.0  
2:00 184.1 F 3.7  
4:00 201.7 F 8.6  
6:00 212.6 F 10.7  
8:00 228.9 F 8.7  
10:00 237.0 F 11.0  
12:00 241.6 F 12.4  
14:00 241.6 F 11.6  
16:00 255.2 F 18.6  
18:00 266.9 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.5 F 29.8  
2:00 271.9 F 29.0  
3:00 271.7 F 29.5  
4:00 272.5 F 29.7  
  
Min 271.3 F 28.9  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.5 F 0.0  
5:00 220.5 F 0.0  
10:00 222.8 F 0.0  
15:00 217.6 F 0.0  
20:00 216.8 F 0.0  
25:00 216.7 F 0.0  
30:00 214.7 F 0.0  
Drying Complete

FILLING: 1:54  
HEATING: 19:28  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 00:58:47  
CYCLE COMPLETE

10/19/2023 02:35 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_