
Midmark M11 - v1.0.0
Total Cycles: 3319

Name: RITTER 1
Sterilizer ID: V2178198
06/01/2023 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.4 F 0.0
2:00 190.9 F 3.5
4:00 209.9 F 9.3
6:00 228.7 F 7.6
8:00 235.2 F 10.9
10:00 234.4 F 8.3
12:00 240.3 F 10.0
14:00 248.2 F 14.2
16:00 261.0 F 21.3
18:00 272.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.7 F 28.9
2:00 271.5 F 28.1
3:00 272.6 F 29.1
4:00 272.4 F 28.6

Min 271.3 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 224.4 F 0.0
10:00 233.2 F 0.0
15:00 228.6 F 0.0
20:00 227.2 F 0.0
25:00 227.8 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 18:29
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:42
CYCLE COMPLETE

06/01/2023 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____