

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1109

Name: RITTER 1  
Sterilizer ID: V2178198  
08/23/2021 09:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.9 F 0.0  
2:00 180.6 F 3.3  
4:00 203.2 F 8.0  
6:00 229.3 F 11.0  
8:00 233.1 F 9.0  
10:00 238.2 F 10.3  
12:00 243.3 F 12.2  
14:00 241.2 F 10.8  
16:00 254.5 F 17.5  
18:00 265.9 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.1 F 28.4  
2:00 272.4 F 28.8  
3:00 271.1 F 28.3  
4:00 272.7 F 28.9  
  
Min 270.9 F 28.1  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.9 F 0.0  
5:00 225.4 F 0.0  
10:00 221.5 F 0.0  
15:00 218.6 F 0.0  
20:00 209.5 F 0.0  
25:00 208.8 F 0.0  
30:00 213.7 F 0.0  
  
Drying Complete

FILLING: 2:35  
HEATING: 19:36  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 00:59:16  
CYCLE COMPLETE

08/23/2021 11:16 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02