
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001301
15:48 17/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:52
PRESSURIZING 4:52
STERILIZING 6:19
136.4 C 223kPa 6:19
136.4 C 226kPa 6:49
136.4 C 225kPa 7:19
136.4 C 225kPa 7:49
136.2 C 222kPa 8:19
136.4 C 225kPa 8:49
136.4 C 222kPa 9:19
136.3 C 224kPa 9:49
136.3 C 224kPa 9:49
VENTING 9:50
AIR DRYING 10:47
CYCLE COMPLETE 25:47

Digital Signature #
97F64E3EEFAA863A
