

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2336

Name: RITTER2  
Sterilizer ID: V2313228  
05/04/2023 07:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.7 F 0.0  
2:00 79.0 F 0.0  
4:00 85.8 F 0.0  
6:00 105.5 F 0.0  
8:00 149.4 F 1.1  
10:00 181.3 F 4.2  
12:00 199.2 F 8.0  
14:00 206.8 F 12.0  
16:00 219.0 F 5.5  
18:00 227.6 F 9.7  
20:00 235.8 F 10.8  
22:00 239.2 F 11.3  
24:00 236.7 F 9.4  
26:00 249.5 F 15.5  
28:00 260.3 F 21.5  
30:00 270.2 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.2 F 29.1  
2:00 272.3 F 29.5  
3:00 271.7 F 29.4  
4:00 271.9 F 29.0

Min 271.4 F 28.8  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.0 F 0.0  
5:00 207.1 F 0.0  
10:00 189.6 F 0.0  
15:00 187.9 F 0.0  
20:00 186.6 F 0.0  
25:00 185.4 F 0.0  
30:00 184.0 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:48  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:08:51  
CYCLE COMPLETE

05/04/2023 08:36 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_