
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000751
9:37 27/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:49
STERILIZING 4:00
136.0 C 220kPa 4:00
136.2 C 231kPa 4:30
136.4 C 233kPa 5:00
136.2 C 231kPa 5:30
136.3 C 232kPa 6:00
136.2 C 230kPa 6:30
136.3 C 232kPa 7:00
136.4 C 232kPa 7:30
136.4 C 232kPa 7:30
VENTING 7:30
AIR DRYING 8:24
CYCLE COMPLETE 8:26

Digital Signature #
EE08523C178D26FC
