

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001952  
7:49 07/07/2021

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:04  
PRESSURIZING 3:07  
STERILIZING 4:12  
136.5 C 222kPa 4:12  
136.1 C 221kPa 4:42  
136.4 C 220kPa 5:12  
136.4 C 224kPa 5:42  
136.1 C 222kPa 6:12  
136.3 C 222kPa 6:42  
136.4 C 221kPa 7:12  
136.3 C 223kPa 7:42  
136.3 C 223kPa 7:42  
VENTING 7:42  
AIR DRYING 8:32  
CYCLE COMPLETE 8:36

-----  
Digital Signature #  
D658AD8BF65C9AB8  
-----