
Midmark M11 - v1.0.0
Total Cycles: 758

Name: RITTER 1
Sterilizer ID: V2178198
03/31/2021 08:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.6 F 0.0
2:00 128.4 F 0.5
4:00 131.5 F 2.3
6:00 153.5 F 5.8
8:00 184.8 F 10.3
10:00 215.3 F 4.5
12:00 222.6 F 8.7
14:00 234.3 F 12.2
16:00 240.5 F 10.7
18:00 235.7 F 8.5
20:00 251.0 F 15.6
22:00 263.3 F 22.7
24:00 270.9 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 271.8 F 28.5
2:00 272.4 F 28.8
3:00 272.8 F 29.1
4:00 271.9 F 29.0

Min 271.4 F 28.4
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.7 F 0.0
5:00 239.0 F 0.0
10:00 228.8 F 0.0
15:00 198.7 F 0.0
20:00 194.4 F 0.0
25:00 181.8 F 0.0
30:00 180.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:44
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:02:40
CYCLE COMPLETE

03/31/2021 09:12 AM
ACCEPTED: X
REJECTED:
Operator: U01