
Midmark M11 - v1.0.0
Total Cycles: 4435

Name: RITTER 1
Sterilizer ID: V2178198
04/08/2024 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.8 F 0.0
2:00 125.1 F 1.1
4:00 137.5 F 3.7
6:00 162.3 F 7.0
8:00 185.4 F 10.9
10:00 206.2 F 3.5
12:00 215.6 F 6.8
14:00 228.0 F 11.1
16:00 228.6 F 6.6
18:00 241.0 F 12.2
20:00 239.0 F 10.0
22:00 251.3 F 16.0
24:00 262.1 F 22.2
26:00 272.2 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.1
2:00 271.7 F 28.4
3:00 272.1 F 29.1
4:00 272.3 F 28.8

Min 271.3 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 209.4 F 0.0
10:00 208.8 F 0.0
15:00 213.5 F 0.0
20:00 223.4 F 0.0
25:00 230.4 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 26:31
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:05:04
CYCLE COMPLETE

04/08/2024 02:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____