

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4057

Name: RITTER 1  
Sterilizer ID: V2178198  
11/27/2023 08:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.3 F 0.0  
2:00 68.7 F 0.2  
4:00 79.0 F 1.0  
6:00 109.8 F 2.8  
8:00 144.6 F 5.1  
10:00 167.7 F 7.8  
12:00 181.5 F 10.7  
14:00 209.1 F 8.3  
16:00 212.9 F 4.7  
18:00 221.8 F 7.9  
20:00 230.4 F 11.5  
22:00 228.0 F 6.4  
24:00 241.3 F 11.6  
26:00 236.5 F 9.0  
28:00 249.1 F 14.8  
30:00 259.5 F 20.6  
32:00 269.3 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 271.7 F 29.0  
2:00 271.8 F 29.1  
3:00 271.8 F 29.2  
4:00 271.7 F 29.1  
  
Min 271.5 F 28.7  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 164.7 F 0.0  
5:00 234.8 F 0.0  
10:00 234.0 F 0.0  
15:00 224.8 F 0.0  
20:00 195.0 F 0.0  
25:00 206.9 F 0.0  
30:00 203.8 F 0.0  
  
Drying Complete

FILLING: 0:53  
HEATING: 33:42  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:12:16  
CYCLE COMPLETE

11/27/2023 10:08 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_