

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003450  
11:28 14/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:00  
PRESSURIZING 2:04  
STERILIZING 2:56  
136.0 C 220kPa 2:56  
136.4 C 225kPa 3:26  
136.4 C 223kPa 3:56  
136.1 C 218kPa 4:26  
136.2 C 222kPa 4:56  
136.5 C 225kPa 5:26  
136.4 C 224kPa 5:56  
136.2 C 223kPa 6:26  
136.2 C 223kPa 6:26  
VENTING 6:26  
AIR DRYING 7:16  
CYCLE COMPLETE 9:13

-----  
Digital Signature #  
910AC3EC5BBF4CF6  
-----