
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004992
10:31 24/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:05
STERILIZING 3:02
136.4 C 222kPa 3:02
136.4 C 220kPa 3:32
135.8 C 219kPa 4:02
136.1 C 220kPa 4:32
136.4 C 221kPa 5:02
136.4 C 221kPa 5:32
136.4 C 221kPa 6:02
136.4 C 224kPa 6:32
136.4 C 224kPa 6:32
VENTING 6:32
AIR DRYING 7:36
CYCLE COMPLETE 10:05

Digital Signature #
0569CD7383F053FB
