
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006209
10:23 30/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:49
PRESSURIZING 1:55
STERILIZING 2:59
136.4 C 220kPa 2:59
136.2 C 223kPa 3:29
136.4 C 224kPa 3:59
136.4 C 224kPa 4:29
136.4 C 222kPa 4:59
136.3 C 224kPa 5:29
136.3 C 224kPa 5:59
136.2 C 223kPa 6:29
136.2 C 223kPa 6:29
VENTING 6:29
AIR DRYING 7:19
CYCLE COMPLETE 29:25

Digital Signature #
7E311F80841DE1F0
