
Midmark M11 - v1.0.0
Total Cycles: 3248

Name: RITTER 1
Sterilizer ID: V2178198
05/17/2023 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.1 F 0.0
2:00 190.6 F 0.0
4:00 191.5 F 2.7
6:00 211.4 F 7.6
8:00 225.1 F 12.8
10:00 230.4 F 8.3
12:00 241.3 F 12.0
14:00 242.0 F 11.7
16:00 239.8 F 10.3
18:00 253.9 F 17.2
20:00 265.7 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.5 F 28.8
2:00 271.7 F 28.1
3:00 272.1 F 28.7
4:00 272.4 F 28.7

Min 271.3 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 217.8 F 0.0
10:00 237.6 F 0.0
15:00 227.6 F 0.0
20:00 222.3 F 0.0
25:00 219.5 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 21:44
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:06
CYCLE COMPLETE

05/17/2023 09:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____