
Midmark M11 - v1.0.0
Total Cycles: 1484

Name: RITTER 1
Sterilizer ID: V2178198
12/29/2021 07:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.4 F 0.0
2:00 75.0 F 0.0
4:00 91.9 F 0.0
6:00 116.6 F 0.2
8:00 145.6 F 1.4
10:00 170.5 F 3.2
12:00 189.1 F 5.4
14:00 203.3 F 8.2
16:00 215.4 F 11.5
18:00 218.9 F 4.6
20:00 231.0 F 9.1
22:00 241.1 F 13.5
24:00 240.0 F 10.9
26:00 236.6 F 8.9
28:00 250.6 F 15.5
30:00 262.0 F 22.0
32:00 272.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.0
2:00 271.8 F 28.2
3:00 272.5 F 29.1
4:00 272.4 F 28.7

Min 271.5 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 167.7 F 0.0
5:00 201.1 F 0.0
10:00 210.7 F 0.0
15:00 205.7 F 0.0
20:00 203.1 F 0.0
25:00 203.3 F 0.0
30:00 206.0 F 0.0

Drying Complete

FILLING: 0:56
HEATING: 32:28
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:10:32
CYCLE COMPLETE

12/29/2021 09:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____