
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005006
9:17 30/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:52
STERILIZING 3:54
136.4 C 223kPa 3:54
136.1 C 221kPa 4:24
136.4 C 221kPa 4:54
136.2 C 223kPa 5:24
136.4 C 222kPa 5:54
136.4 C 221kPa 6:24
136.4 C 222kPa 6:54
136.3 C 221kPa 7:24
136.3 C 221kPa 7:24
VENTING 7:24
AIR DRYING 8:10
CYCLE COMPLETE 13:47

Digital Signature #
E3447E2C2991AB84
