

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002100  
12:10 01/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:09  
PRESSURIZING 2:15  
STERILIZING 3:15  
136.4 C 224kPa 3:15  
136.3 C 224kPa 3:45  
136.3 C 225kPa 4:15  
136.4 C 222kPa 4:45  
136.4 C 223kPa 5:15  
136.4 C 222kPa 5:45  
136.4 C 225kPa 6:15  
136.4 C 225kPa 6:45  
136.4 C 225kPa 6:45  
VENTING 6:45  
AIR DRYING 7:37  
CYCLE COMPLETE 37:38  
-----

Digital Signature #  
C6E8CA2B65681B98  
-----