

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3847

Name: RITTER2  
Sterilizer ID: V2313228  
08/12/2024 11:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.7 F 0.0  
2:00 167.2 F 2.3  
4:00 184.2 F 6.1  
6:00 218.2 F 11.1  
8:00 221.4 F 6.4  
10:00 231.9 F 11.8  
12:00 234.0 F 9.3  
14:00 237.4 F 9.6  
16:00 248.0 F 14.7  
18:00 259.4 F 21.4  
20:00 270.3 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.5 F 29.5  
1:00 272.1 F 29.6  
2:00 272.0 F 29.6  
3:00 272.0 F 29.7  
4:00 272.1 F 29.9  
  
Min 271.2 F 29.4  
Max 272.6 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.1 F 0.0  
5:00 205.9 F 0.0  
10:00 193.7 F 0.0  
15:00 191.0 F 0.0  
20:00 189.5 F 0.0  
25:00 187.6 F 0.0  
30:00 185.2 F 0.0  
Drying Complete

FILLING: 1:30  
HEATING: 20:39  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 00:59:38  
CYCLE COMPLETE

08/12/2024 12:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_