
Midmark M11 - v1.0.0
Total Cycles: 1195

Name: RITTER 1
Sterilizer ID: V2178198
09/22/2021 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.1 F 0.0
2:00 187.1 F 3.4
4:00 200.8 F 8.0
6:00 218.6 F 12.9
8:00 228.0 F 7.2
10:00 239.5 F 12.5
12:00 239.6 F 10.5
14:00 237.5 F 9.2
16:00 251.9 F 16.2
18:00 263.6 F 22.9
20:00 271.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.8
2:00 271.6 F 28.1
3:00 272.5 F 29.0
4:00 272.2 F 28.5

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 200.0 F 0.0
10:00 193.7 F 0.0
15:00 190.0 F 0.0
20:00 204.8 F 0.0
25:00 229.1 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 20:07
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 00:59:37
CYCLE COMPLETE

09/22/2021 12:17 PM
ACCEPTED: X
REJECTED:
Operator: U01