
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000100
10:48 02/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:13
PRESSURIZING 4:13
STERILIZING 5:35
136.5 C 226kPa 5:35
136.4 C 226kPa 6:05
136.2 C 224kPa 6:35
136.3 C 225kPa 7:05
136.4 C 227kPa 7:35
136.2 C 225kPa 8:05
136.3 C 225kPa 8:35
136.1 C 224kPa 9:05
136.1 C 224kPa 9:05
VENTING 9:05
AIR DRYING 10:07
CYCLE COMPLETE 25:07

Digital Signature #
6378AB574F2B2D0F
