
Midmark M11 - v1.0.0
Total Cycles: 2742

Name: RITTER 1
Sterilizer ID: V2178198
01/05/2023 12:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.7 F 0.0
2:00 190.3 F 0.0
4:00 196.8 F 1.9
6:00 209.6 F 6.4
8:00 222.1 F 11.3
10:00 226.4 F 6.7
12:00 239.1 F 12.1
14:00 237.1 F 9.6
16:00 237.5 F 8.9
18:00 248.9 F 14.7
20:00 260.9 F 21.4
22:00 271.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.6 F 28.9
2:00 271.5 F 28.1
3:00 272.7 F 29.0
4:00 271.7 F 28.3

Min 271.3 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 219.1 F 0.0
10:00 238.9 F 0.0
15:00 235.2 F 0.0
20:00 228.0 F 0.0
25:00 227.2 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 22:37
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:38
CYCLE COMPLETE

01/05/2023 01:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____