
Midmark M11 - v1.0.0
Total Cycles: 2575

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2022 11:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.5 F 0.0
2:00 183.7 F 4.2
4:00 194.7 F 9.1
6:00 215.0 F 7.4
8:00 229.0 F 9.0
10:00 232.9 F 9.7
12:00 241.0 F 11.9
14:00 239.9 F 10.4
16:00 254.0 F 17.5
18:00 265.9 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.1 F 28.8
2:00 271.5 F 28.8
3:00 271.9 F 28.8
4:00 271.3 F 28.6

Min 271.2 F 28.3
Max 272.3 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 221.8 F 0.0
10:00 238.3 F 0.0
15:00 234.3 F 0.0
20:00 228.0 F 0.0
25:00 229.9 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:49
CYCLE COMPLETE

11/17/2022 12:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____