

-----  
Midmark M11 - v1.0.5  
Total Cycles: 255

Name: RITTER2  
Sterilizer ID: V2313228  
07/28/2021 07:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.7 F 0.0  
2:00 158.0 F 1.6  
4:00 159.4 F 4.8  
6:00 177.9 F 8.7  
8:00 201.2 F 13.2  
10:00 216.6 F 7.2  
12:00 230.4 F 12.1  
14:00 232.7 F 9.1  
16:00 238.6 F 10.3  
18:00 245.4 F 13.6  
20:00 257.1 F 19.8  
22:00 267.7 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.2  
1:00 272.7 F 29.8  
2:00 271.7 F 29.1  
3:00 272.7 F 30.0  
4:00 272.1 F 29.4

Min 271.7 F 29.1  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.9 F 0.0  
5:00 222.6 F 0.0  
10:00 223.3 F 0.0  
15:00 204.4 F 0.0  
20:00 225.7 F 0.0  
25:00 242.4 F 0.0  
30:00 173.5 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 23:23  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:01:42  
CYCLE COMPLETE

07/28/2021 08:37 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01