
Midmark M11 - v1.0.0
Total Cycles: 2172

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2022 04:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.6 F 0.0
2:00 180.6 F 4.0
4:00 191.6 F 8.7
6:00 209.4 F 11.1
8:00 222.0 F 8.1
10:00 234.8 F 13.1
12:00 236.7 F 10.1
14:00 235.8 F 8.3
16:00 248.2 F 14.5
18:00 259.9 F 21.0
20:00 270.4 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.9 F 29.6
2:00 272.2 F 29.0
3:00 271.9 F 29.2
4:00 272.8 F 29.6

Min 271.6 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 215.5 F 0.0
10:00 240.1 F 0.0
15:00 239.0 F 0.0
20:00 235.0 F 0.0
25:00 229.8 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:12
CYCLE COMPLETE

07/25/2022 05:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____