
Midmark M11 - v1.0.0
Total Cycles: 4080

Name: RITTER 1
Sterilizer ID: V2178198
12/04/2023 03:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.0 F 0.0
2:00 122.1 F 1.1
4:00 164.9 F 3.7
6:00 180.4 F 7.1
8:00 199.6 F 11.0
10:00 221.7 F 4.5
12:00 226.2 F 7.8
14:00 235.6 F 11.8
16:00 232.3 F 7.7
18:00 243.4 F 12.6
20:00 238.8 F 10.3
22:00 250.2 F 15.8
24:00 260.4 F 21.6
26:00 270.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.9 F 29.6
2:00 271.6 F 28.9
3:00 272.4 F 29.5
4:00 272.5 F 29.4

Min 271.3 F 28.8
Max 273.4 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.8 F 0.0
5:00 242.7 F 0.0
10:00 205.8 F 0.0
15:00 208.8 F 0.0
20:00 203.1 F 0.0
25:00 205.0 F 0.0
30:00 208.7 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 26:57
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:05:53
CYCLE COMPLETE

12/04/2023 05:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____