
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002656
11:56 07/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:47
STERILIZING 3:57
136.0 C 221kPa 3:57
136.4 C 226kPa 4:27
136.3 C 225kPa 4:57
136.1 C 222kPa 5:27
136.1 C 223kPa 5:57
136.0 C 222kPa 6:27
136.2 C 224kPa 6:57
136.2 C 223kPa 7:27
136.2 C 223kPa 7:27
VENTING 7:27
AIR DRYING 8:20
CYCLE COMPLETE 38:20

Digital Signature #
2F4678071CE6774E
