
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006760
9:17 19/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:22
PRESSURIZING 3:22
STERILIZING 4:30
136.4 C 223kPa 4:30
136.4 C 223kPa 5:00
136.4 C 223kPa 5:30
136.4 C 224kPa 6:00
135.9 C 221kPa 6:30
136.4 C 221kPa 7:00
136.1 C 221kPa 7:30
136.4 C 221kPa 8:00
136.4 C 221kPa 8:00
VENTING 8:00
AIR DRYING 8:52
CYCLE COMPLETE 9:10

Digital Signature #
12319D2C6631E0A9
