

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1604

Name: RITTER2  
Sterilizer ID: V2313228  
10/17/2022 08:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.4 F 0.0  
2:00 156.2 F 1.9  
4:00 156.8 F 4.9  
6:00 179.9 F 9.1  
8:00 203.1 F 8.8  
10:00 219.6 F 8.2  
12:00 232.0 F 13.1  
14:00 234.2 F 9.9  
16:00 236.0 F 9.0  
18:00 247.0 F 14.4  
20:00 258.7 F 20.8  
22:00 269.3 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.3  
1:00 272.4 F 30.1  
2:00 272.1 F 29.6  
3:00 271.2 F 29.2  
4:00 272.0 F 30.0

Min 271.2 F 29.1  
Max 272.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.9 F 0.0  
5:00 207.7 F 0.0  
10:00 216.8 F 0.0  
15:00 220.0 F 0.0  
20:00 218.7 F 0.0  
25:00 218.8 F 0.0  
30:00 220.3 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 23:11  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:01:30  
CYCLE COMPLETE

10/17/2022 09:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_