
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002303
9:46 27/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 2:59
STERILIZING 3:59
136.4 C 221kPa 3:59
136.3 C 224kPa 4:29
136.1 C 222kPa 4:59
136.4 C 223kPa 5:29
136.3 C 224kPa 5:59
136.4 C 224kPa 6:29
136.2 C 220kPa 6:59
136.4 C 223kPa 7:29
136.4 C 223kPa 7:29
VENTING 7:29
AIR DRYING 8:22
CYCLE COMPLETE 8:27

Digital Signature #
95B67AC386741722
