

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2773

Name: RITTER2  
Sterilizer ID: V2313228  
08/15/2023 04:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.2 F 0.0  
2:00 201.1 F 2.1  
4:00 217.3 F 8.0  
6:00 229.8 F 12.1  
8:00 235.8 F 10.8  
10:00 235.8 F 9.2  
12:00 238.1 F 9.4  
14:00 250.5 F 15.9  
16:00 262.9 F 23.0  
18:00 272.0 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.3 F 29.1  
2:00 272.5 F 29.7  
3:00 271.7 F 28.7  
4:00 272.7 F 29.4  
  
Min 271.5 F 28.6  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.0 F 0.0  
5:00 223.6 F 0.0  
10:00 230.0 F 0.0  
15:00 225.3 F 0.0  
20:00 223.8 F 0.0  
25:00 222.1 F 0.0  
30:00 220.6 F 0.0  
Drying Complete

FILLING: 1:60  
HEATING: 18:07  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:22  
CYCLE COMPLETE

08/15/2023 05:11 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_