
Midmark M11 - v1.0.5
Total Cycles: 3895

Name: RITTER2
Sterilizer ID: V2313228
08/27/2024 12:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.8 F 0.0
2:00 180.9 F 2.4
4:00 180.3 F 6.5
6:00 204.5 F 12.0
8:00 220.3 F 6.6
10:00 235.0 F 12.2
12:00 236.6 F 9.8
14:00 237.0 F 8.9
16:00 250.0 F 15.5
18:00 262.1 F 22.4
20:00 272.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.1 F 28.8
2:00 272.7 F 29.2
3:00 272.2 F 29.3
4:00 271.9 F 28.6

Min 271.7 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 187.6 F 0.0
10:00 219.7 F 0.0
15:00 220.8 F 0.0
20:00 220.9 F 0.0
25:00 219.4 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 20:14
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:59
CYCLE COMPLETE

08/27/2024 01:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____