

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 005056  
12:59 02/02/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:26  
PRESSURIZING 2:30  
STERILIZING 3:21  
136.0 C 220kPa 3:21  
136.2 C 222kPa 3:51  
136.4 C 221kPa 4:21  
136.1 C 222kPa 4:51  
136.4 C 224kPa 5:21  
136.3 C 224kPa 5:51  
136.4 C 223kPa 6:21  
136.3 C 222kPa 6:51  
136.3 C 222kPa 6:51  
VENTING 6:51  
AIR DRYING 7:54  
CYCLE COMPLETE 17:40

-----  
Digital Signature #  
35732610D1251AA2  
-----