
Midmark M11 - v1.0.5
Total Cycles: 2057

Name: RITTER2
Sterilizer ID: V2313228
02/15/2023 02:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.2 F 0.0
2:00 192.5 F 3.6
4:00 204.8 F 8.5
6:00 222.8 F 10.0
8:00 226.2 F 9.9
10:00 232.4 F 8.4
12:00 242.8 F 12.5
14:00 244.8 F 13.1
16:00 257.7 F 20.0
18:00 268.9 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.1 F 29.5
2:00 271.3 F 29.1
3:00 272.1 F 29.5
4:00 271.3 F 28.9

Min 271.2 F 28.9
Max 272.2 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.4 F 0.0
5:00 228.3 F 0.0
10:00 230.0 F 0.0
15:00 228.0 F 0.0
20:00 225.3 F 0.0
25:00 225.4 F 0.0
30:00 227.6 F 0.0

Drying Complete

FILLING: 1:47
HEATING: 19:05
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

02/15/2023 03:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____