

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3681

Name: RITTER2  
Sterilizer ID: V2313228  
06/10/2024 11:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 102.6 F 0.0  
2:00 104.1 F 0.5  
4:00 119.3 F 2.4  
6:00 145.9 F 5.1  
8:00 170.8 F 8.2  
10:00 189.6 F 11.6  
12:00 210.9 F 4.1  
14:00 214.4 F 7.5  
16:00 226.0 F 11.3  
18:00 226.9 F 7.1  
20:00 237.9 F 12.2  
22:00 237.3 F 10.2  
24:00 249.0 F 15.9  
26:00 259.7 F 21.7  
28:00 269.2 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.6  
1:00 272.1 F 30.3  
2:00 272.5 F 30.2  
3:00 272.1 F 29.8  
4:00 271.3 F 29.4

Min 271.2 F 29.4  
Max 272.7 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.0 F 0.0  
5:00 205.7 F 0.0  
10:00 201.4 F 0.0  
15:00 194.8 F 0.0  
20:00 195.1 F 0.0  
25:00 196.8 F 0.0  
30:00 197.3 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 29:13  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:07:48  
CYCLE COMPLETE

06/10/2024 12:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_