

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1835

Name: RITTER2  
Sterilizer ID: V2313228  
12/13/2022 05:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 137.3 F 0.0  
2:00 138.7 F 0.0  
4:00 157.8 F 0.7  
6:00 186.8 F 5.3  
8:00 214.2 F 10.6  
10:00 223.4 F 6.1  
12:00 237.2 F 12.0  
14:00 236.2 F 9.8  
16:00 237.7 F 8.8  
18:00 250.1 F 15.8  
20:00 262.7 F 23.0  
22:00 272.1 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.4 F 29.5  
2:00 271.2 F 28.9  
3:00 272.3 F 29.7  
4:00 271.7 F 29.1

Min 271.2 F 28.8  
Max 272.4 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.5 F 0.0  
5:00 198.9 F 0.0  
10:00 233.2 F 0.0  
15:00 233.5 F 0.0  
20:00 226.6 F 0.0  
25:00 225.4 F 0.0  
30:00 226.7 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 22:11  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:00:39  
CYCLE COMPLETE

12/13/2022 06:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_