
Midmark M11 - v1.0.0
Total Cycles: 3392

Name: RITTER 1
Sterilizer ID: V2178198
06/15/2023 07:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.3 F 0.0
2:00 81.3 F 0.0
4:00 94.2 F 0.0
6:00 111.3 F 0.1
8:00 143.6 F 2.5
10:00 168.3 F 5.4
12:00 189.6 F 8.8
14:00 208.2 F 12.6
16:00 217.0 F 5.7
18:00 227.7 F 9.7
20:00 235.0 F 11.0
22:00 237.5 F 10.4
24:00 235.3 F 8.1
26:00 247.2 F 14.1
28:00 258.6 F 20.3
30:00 268.9 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 29.3
2:00 271.5 F 28.4
3:00 272.4 F 29.4
4:00 272.2 F 28.8

Min 271.3 F 28.4
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 214.2 F 0.0
10:00 207.4 F 0.0
15:00 197.9 F 0.0
20:00 222.6 F 0.0
25:00 226.9 F 0.0
30:00 232.1 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:10
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:09:11
CYCLE COMPLETE

06/15/2023 08:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____