
Midmark M11 - v1.0.5
Total Cycles: 2320

Name: RITTER2
Sterilizer ID: V2313228
05/01/2023 12:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.4 F 0.0
2:00 198.8 F 4.1
4:00 200.8 F 9.3
6:00 221.0 F 7.8
8:00 232.2 F 11.0
10:00 233.3 F 9.0
12:00 237.9 F 9.8
14:00 248.2 F 15.2
16:00 260.9 F 22.3
18:00 272.2 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 272.4 F 29.9
2:00 271.6 F 29.7
3:00 272.4 F 29.9
4:00 271.5 F 29.4

Min 271.4 F 29.3
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 227.6 F 0.0
10:00 237.4 F 0.0
15:00 230.3 F 0.0
20:00 225.3 F 0.0
25:00 228.1 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 2:07
HEATING: 18:28
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:57:54
CYCLE COMPLETE

05/01/2023 01:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____