

-----  
Midmark M11 - v1.0.0  
Total Cycles: 761

Name: RITTER 1  
Sterilizer ID: V2178198  
03/31/2021 12:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 120.7 F 0.0  
2:00 128.4 F 0.9  
4:00 147.3 F 3.8  
6:00 177.6 F 7.8  
8:00 201.1 F 12.8  
10:00 214.7 F 6.2  
12:00 228.3 F 11.3  
14:00 233.8 F 7.7  
16:00 245.0 F 12.3  
18:00 244.2 F 12.2  
20:00 257.1 F 19.0  
22:00 268.6 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.6  
1:00 272.2 F 29.0  
2:00 272.6 F 29.0  
3:00 272.5 F 28.9  
4:00 272.5 F 28.8

Min 271.4 F 28.5  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.6 F 0.0  
5:00 240.4 F 0.0  
10:00 230.8 F 0.0  
15:00 202.3 F 0.0  
20:00 218.9 F 0.0  
25:00 218.4 F 0.0  
30:00 219.1 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 23:39  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 30:00  
TOTAL CYCLE: 01:01:37  
CYCLE COMPLETE

03/31/2021 01:18 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01