
Midmark M11 - v1.0.5
Total Cycles: 2263

Name: RITTER2
Sterilizer ID: V2313228
04/18/2023 02:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.0 F 0.0
2:00 189.4 F 4.3
4:00 206.3 F 9.9
6:00 221.4 F 6.6
8:00 236.4 F 11.7
10:00 236.3 F 9.9
12:00 235.7 F 8.8
14:00 250.6 F 16.2
16:00 263.1 F 23.4
18:00 271.9 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.2 F 29.3
2:00 272.5 F 29.8
3:00 271.6 F 29.1
4:00 272.5 F 29.7

Min 271.6 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 220.9 F 0.0
10:00 233.5 F 0.0
15:00 240.2 F 0.0
20:00 233.3 F 0.0
25:00 224.9 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 18:04
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:33
CYCLE COMPLETE

04/18/2023 03:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____