

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000978  
9:24 10/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:15  
PRESSURIZING 2:25  
STERILIZING 3:27  
136.0 C 228kPa 3:27  
136.4 C 232kPa 3:57  
136.2 C 230kPa 4:27  
136.4 C 232kPa 4:57  
136.4 C 231kPa 5:27  
136.4 C 231kPa 5:57  
136.4 C 232kPa 6:27  
136.4 C 233kPa 6:57  
136.4 C 233kPa 6:57  
VENTING 6:57  
CYCLE ABORTED 7:46  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 7:46  
-----

Digital Signature #  
81683CCD8997C70C  
-----