

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5150

Name: RITTER 1  
Sterilizer ID: V2178198  
12/03/2024 11:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.4 F 0.0  
2:00 168.9 F 0.2  
4:00 173.4 F 0.5  
6:00 197.1 F 1.3  
8:00 211.0 F 5.4  
10:00 227.0 F 9.6  
12:00 234.5 F 10.0  
14:00 237.7 F 11.1  
16:00 236.1 F 8.9  
18:00 239.9 F 9.7  
20:00 249.1 F 14.6  
22:00 261.0 F 21.2  
24:00 271.3 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.0  
1:00 272.1 F 28.7  
2:00 272.4 F 28.7  
3:00 272.0 F 28.3  
4:00 271.3 F 27.9

Min 271.2 F 27.9  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.2 F 0.0  
5:00 207.2 F 0.0  
10:00 217.9 F 0.0  
15:00 234.1 F 0.0  
20:00 230.2 F 0.0  
25:00 228.5 F 0.0  
30:00 227.3 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 24:47  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:03:20  
CYCLE COMPLETE

12/03/2024 12:22 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_