

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007

Unit # : 000

WATER QUALITY

0.2uS / 0.1ppm

CYCLE NUMBER 001457

10:56 18/09/2023

WRAPPED

135 C for 10 min.

CYCLE START 0:00

CONDITIONING 2:27

PRESSURIZING 4:25

STERILIZING 5:25

136.4 C 224kPa 5:25

136.2 C 224kPa 5:55

136.3 C 225kPa 6:25

136.2 C 223kPa 6:55

136.1 C 222kPa 7:25

136.2 C 224kPa 7:55

136.2 C 223kPa 8:25

136.4 C 225kPa 8:55

136.1 C 223kPa 9:25

136.4 C 226kPa 9:55

136.2 C 223kPa 10:25

136.4 C 226kPa 10:55

136.4 C 225kPa 11:25

136.1 C 223kPa 11:55

136.4 C 226kPa 12:25

136.4 C 225kPa 12:55

136.2 C 223kPa 13:25

136.4 C 226kPa 13:55

136.4 C 226kPa 14:25

136.4 C 223kPa 14:55

136.2 C 224kPa 15:25

136.2 C 224kPa 15:25

VENTING 15:25

AIR DRYING 16:15

CYCLE COMPLETE 17:50  
-----

Digital Signature #  
647D513A56395B4F  
-----